

To determine the moisture content and regain percentage of textile materials. Drying is performed by circulation of warm air electrically heated, regulated by an electronic temperature controller (range $0^{\circ}/180^{\circ}$ C, accuracy $\pm 1^{\circ}$ C). The material to be tested is dried within a metal container suspended from an electronic balance. Two successive weights of the same value determine the point at which moisture content can be calculated. Includes an electronic balance with 2100 g capacity and 0.01 g accuracy. The sample is weighed automatically by means of a basket connected to the balance.

SIZE

WEIGHT	130 kg / 286 lbs	
DIMENSIONS	1000 mm (w) x 700 mm (d) x 1280 mm (h) / 40 in (w) x 28 in (d) x 54 in (h)	

ORDERING INFORMATION

300064 Moi:

STANDARDS:

ASTM D2564	BS 4784	ISO 6741
------------	---------	----------

Fibers